

JVCKENWOOD

EU Declaration of Conformity

No.: 10270 L - 17027HEA.029

Model name: CL19196-xxxx **Brand:** JVC

Batch: Serial number(s):

Description: LCD Monitor CCL196
Medical Electrical Equipment LCD Monitor
class I

Medical device class: class I
Accessory: Digital signal cable (DVI-D - DVI-D), Analog signal cable (VGA - VGA)

Manufacturer: JVC KENWOOD Corporation
3-12, Moriya-cho, Kanagawa-ku,
Yokohama-shi, Kanagawa, 221-0022, JAPAN

Authorised Representative: JVCKENWOOD Deutschland GmbH
Konrad-Adenauer-Allee 1-11, 61118 Bad Vilbel, GERMANY

This declaration of conformity is issued under the sole responsibility of the manufacturer.

The object of the declaration described above is in conformity with the relevant Union harmonisation legislation:

Directive 93/42/EEC (medical devices)

harmonised standards and technical specifications: electromagnetic compatibility:
EN 60601-1-2:2007 Class B

safety:

EN 60601-1:2006 + A1:2013 + A12:2014

others (usability, medical device software - software life-cycle processes):

EN 62366:2008; EN 60601-1-6:2010 + A1:2015; EN 62304:2006

Directive 2011/65/EU (restriction of the use of certain hazardous substances)

harmonised standards: EN 50581:2012

Notified body (where applicable):

- notified body number: -

Reference number of the certificate of notified body: -

Additional information: The models "CL19196-J00U", "CL19196-J00E", "CL19196-J001", "CL19196-J002", "CL19196-B0S3", "CL19196-B0S4" and "CL19196-J005" are identical to "CL19196-xxxx" in every aspect of attributes. The model numbers above are only used for marketing purposes.

Signed for and on behalf of: JVCKENWOOD Deutschland GmbH

Bad Vilbel, 2017/07/06

Place, Date (YYYY/MM/DD)


Dipl.-Ing. André Overbeck
Product Compliance Engineer

This document may not be changed or copied without authorisation.